



**Office of the Chief Executive Officer
Shri Mata Vaishno Devi Shrine Board, Katra**

Subject: SITC and 5 years post-warranty CAMC for Medical Equipments for Medical College, Kakryal.

Reference: e-NIT No. CO/Pur/NE/612-I/178 dated 19.04.2025

CORRIGENDUM / ADDENDUM

Item Name: (Bubble C-PAP)

| S. No. | Existing Specifications | Revised Specifications |
|---------------|---|--|
| 4 | Humidifier should be USFDA & CF certified | It should be USFDA or EU-CE or BIS / CDSCO / IEC Test report |
| 12. | Max peak flow should be 180 Ltr / min | It should be 15 Ltr / min |
| 21. | Certification: The entire system including Air oxygen blender should be USFDA & CE approved | It should be USFDA / EU-CE/BIS/CDSCO |

Other terms & conditions of the ibid e-NIT shall remain unchanged.

This issues with approval of the Addl. Chief Executive Officer.

**Sd/-
(Dr. Gopal K Sharma)
Asstt. Chief Executive Officer**

**No. CO/Pur/NE/612-I/383
Dated: 06/05/2025**